

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

JULY 24, 2009

OPERATOR'S NAME:

LAWRENSON/BRANDON/J

OPERATOR'S NUMBER: 6609

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0123456

DEPT/AGENCY: JNU1

CASE/REPORT: 0123456

TEST TYPE: V

ALCO TARGET VALUE: .086 ✓

ALCO S/N: X301556

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 30.27 in
ALCO TARGET .087 10:56
BLANK TEST .000 10:57
INTERNAL STANDARD VERIFIED 10:57
ALCO TV 30.27 in .088 ✓ 10:58
BLANK TEST .000 10:59
SUBJECT SAMPLE .000 10:59
BLANK TEST .000 11:00
ALCO TV 30.27 in .088 ✓ 11:00
BLANK TEST .000 11:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

JULY 24, 2009

TIME 10:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!*"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 28 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130185 ✓
Name BRANDON LAWRENSEN ID# 6609 Date 07/24/09 ✓
A Agency JUNEAU POLICE DEPT Phone # 586-0600
Instrument Location JUNEAU POLICE DEPT ✓

B Alco S/N X301556 ✓ Target Value .086 ✓ High Pressure 1100
Alco Test Values .088 ✓ .088 ✓
1st Alco 2nd Alco

Signature [Signature] ✓ CB
9/7/09

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory. (OVER)
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program. (Do Not Write in Program)
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of August, 2009.
[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

